

Rat IgG-heavy and light chain cross-adsorbed Antibody

Goat Polyclonal

Conjugate R-Phycoerythrin

Antigen Affinity Purified

Catalog No. A110-305PE

Lot No. A110-305PE-2



APPLICATIONS	F, IF				
SPECIES REACTIVITY	Rat. Minimum reactivity to bovine, chicken, goat, human, mouse, rabbit and sheep				
ISOTYPE	IgG				
AMOUNT	1 ml at 0.5 mg/ml				
STORAGE/SHELF LIFE	2 - 8° C / 1 year from date of receipt				
PHYSICAL STATE	Liquid				
BUFFER	Phosphate Buffered Saline (PBS) containing 0.2% BSA and 0.09% Sodium Azide				
ORIGIN	USA				
PRODUCTION PROCEDURES	<p>Antiserum was cross adsorbed using bovine, chicken, goat, human, mouse, rabbit and sheep immunosorbents to remove cross reactive antibodies. The antibody to rat IgG was isolated by affinity chromatography using antigen coupled to agarose beads and conjugated to R-phycoerythrin (RPE).</p> <p>Antibody concentration was determined by extinction coefficient: absorbance at 280 nm of 1.4 equals 1.0 mg of IgG.</p> <p>By immunoelectrophoresis and ELISA this antibody reacts specifically with rat IgG and with light chains common to other rat immunoglobulins. No antibody was detected against non-immunoglobulin serum proteins. Less than 1% cross reactivity to bovine, chicken, goat, human, mouse, rabbit and sheep IgG was detected. This antibody may cross react with IgG from other species.</p>				
APPLICATIONS	<p>Centrifuge tube to remove product from lid. Optimal working dilutions should be determined experimentally by the investigator. Prepare working dilution immediately before use.</p> <table><tr><td>Flow Cytometry</td><td>1:50 - 1:200</td></tr><tr><td>Immunofluorescence</td><td>1:50 - 1:200</td></tr></table>	Flow Cytometry	1:50 - 1:200	Immunofluorescence	1:50 - 1:200
Flow Cytometry	1:50 - 1:200				
Immunofluorescence	1:50 - 1:200				
APPLICATION NOTES	Not all listed applications have been specifically tested by our laboratory.				
ADDITIONAL INFO	<p>https://www.bethyl.com/product/A110-305PE</p> <p>Use the link above to view SDS, a current list of citations, and other product specific information.</p>				

This document certifies that this product has met all of the quality control standards defined by Bethyl Laboratories, Inc.

Eric McIntush, PhD | Chief Scientific Officer

Date: December 3, 2018