Human IgA cross-adsorbed Antibody

Human IgA cross-adsorbed Antibody	
Goat Polyclonal Antigen Affinity Puri Catalog No. A80-	Conjugate DyLight® 594
5	202D4-3 BETHYL LABORATORIES, INC
APPLICATIONS	IHC, ICC, F, IF
SPECIES REACTIVITY	Human. Minimum reactivity to mouse and rat
ISOTYPE	IgG
AMOUNT	1 ml at 0.5 mg/ml
STORAGE/SHELF LIFE	2 - 8° C / 1 year from date of receipt
PHYSICAL STATE	Liquid
FLUOROPHORE/PROTE	IN 5.5
BUFFER	Phosphate Buffered Saline (PBS) containing 0.2% BSA and 0.09% Sodium Azide
ORIGIN	USA
PRODUCTION PROCEDURES	Antiserum was solid phase adsorbed to ensure class specificity. Antiserum was cross adsorbed using mouse and rat immunosorbents to remove cross reactive antibodies. The antibody to human IgA was isolated by affinity chromatography using antigen coupled to agarose beads and conjugated to DyLight® 594.
	Antibody concentration was determined by extinction coefficient: absorbance at 280 nm of 1.4 equals 1.0 mg of IgG.
	By immunoelectrophoresis and ELISA this antibody reacts specifically with human IgA. Cross reactivity with IgM and IgG is negligible. No antibody was detected against non-immunoglobulin serum proteins. Less than 1% cross reactivity to mouse and rat IgA was detected. This antibody may cross react with IgA from other species.
APPLICATIONS	Centrifuge tube to remove product from lid. Optimal working dilutions should be determined experimentally by the investigator. Prepare working dilution immediately before use.
	Immunohistochemistry 1:50 – 1:500
	Immunocytochemistry 1:50 – 1:500
	Flow Cytometry 1:50 – 1:200
	Immunofluorescence 1:50 – 1:500
APPLICATION NOTES	Not all listed applications have been specifically tested by our laboratory.
	DyLight® 594 is excited at 593 (in PBS) and emits at 618 (in PBS).
	DyLight® is a trademark of Thermo Fisher Scientific Inc. and its subsidiaries.
ADDITIONAL INFO	https://www.bethyl.com/product/A80-202D4 Use the link above to view SDS, a current list of citations, and other product specific information.

This document certifies that this product has met all of the quality control standards defined by Bethyl Laboratories, Inc. Eric McIntush, PhD | Chief Scientific Officer Date: December 3, 2018

2

Bethyl Laboratories, Inc. • 25043 West FM 1097 • Montgomery, TX 77356 • 800.338.9579 • 936.597.6111 • 866.597.6105 (FAX) • www.bethyl.com • technical@bethyl.com