

# Human IgG Heavy and Light Chain Cross-Adsorbed Antibody

Donkey Polyclonal Conjugate DyLight® 800

Antigen Affinity Purified

Catalog No. A80-246D8

Lot No. 10

---

<b>APPLICATIONS</b>	IHC, ICC, Flow Cyt, IF
<b>SPECIES REACTIVITY</b>	Human. Minimum reactivity to bovine, chicken, goat, mouse, rabbit, rat and sheep
<b>AMOUNT</b>	1 ml
<b>CONCENTRATION</b>	0.5 mg/ml
<b>STORAGE/SHELF LIFE</b>	2 – 8°C / 1 year from date of receipt
<b>PHYSICAL STATE</b>	Liquid
<b>BUFFER</b>	Phosphate Buffered Saline (PBS) containing 0.2% BSA and 0.09% Sodium Azide
<b>FLUOROPHORE/PROTEIN</b>	3.8
<b>ISOTYPE</b>	IgG
<b>ORIGIN</b>	USA
<b>PRODUCTION PROCEDURES</b>	Antiserum was cross adsorbed using bovine, chicken, goat, mouse, rabbit, rat & sheep immunosorbents to remove cross reactive antibodies. The antibody to human IgG was isolated by affinity chromatography using antigen coupled to agarose beads and conjugated to DyLight® 800.

Antibody concentration was determined by extinction coefficient: absorbance at 280 nm of 1.4 equals 1.0 mg of IgG.

By immunoelectrophoresis and ELISA this antibody reacts specifically with human IgG and with light chains common to other human immunoglobulins. No antibody was detected against non-immunoglobulin serum proteins. Less than 1% cross reactivity to bovine, chicken, goat, mouse, rabbit, rat & sheep IgG was detected. This antibody may cross react with IgG from other species.

**APPLICATIONS** Centrifuge tube to remove product from lid. Optimal working dilutions should be determined experimentally by the investigator. Prepare working dilution immediately before use.

Immunohistochemistry	1:50 – 1:500
Immunocytochemistry	1:50 – 1:500
Flow Cytometry	1:50 – 1:200
Immunofluorescence	1:50 – 1:500

**APPLICATION NOTES** Not all listed applications have been specifically tested by our laboratory.

DyLight® 800 is excited at 770 (in PBS) and emits at 794 (in PBS).

DyLight® is a trademark of Thermo Fisher Scientific Inc. and its subsidiaries.

**ADDITIONAL INFO** <https://www.fortislife.com/p/A80-246D8>

Use the link above to view SDS, a current list of citations, and other product specific information.

This document certifies that this product has met all of the quality control standards defined by Bethyl Laboratories, Inc.  
Michael Spencer, PhD Date: May 3, 2023